



OFFICE OF THE  
**Technical Advisory Committee**  
RAWALPINDI INSTITUTE OF CARDIOLOGY  
RAWAL ROAD, RAWALPINDI



No: RIC/TAC/58/2022

Dated: 26/04/2022

**MINUTES OF MEETING OF TECHNICAL ADVISORY COMMITTEE FOR EVALUATION OF ANNUAL TENDER OF SURGICAL & DISPOSABLE ITEMS FOR OT, ANESTHESIA & OTHER DEPARTMENTS FOR FY 2022-23.**

**REF NO. RIC/PO/4466/22, Dated 24-03-2022.**

A Meeting of Technical Advisory Committee for evaluation of Annual Tender of Surgical & Disposable Items For OT, Anesthesia & Other Departments For FY 2022-23 was held on 26/04/2022 at 09:00 am, in the office of Technical Advisory Committee, Rawalpindi Institute of Cardiology. Following members attended the meeting.

**LIST OF MEMBERS:**

S#	Members	Designation
1.	Dr. Javed Iqbal	Chairman, TAC, RIC Rwp.
2.	Dr. Gul-e-Raana Waseem	Member/ Secretary, TAC, RIC, Rwp.
3.	Dr. Humaira Farooq Kiani	Pharmacist, Member, TAC, RIC, Rwp.
4.	M. Noor-ul-Amin	Medical Technologist, Member, TAC, RIC, Rwp.
5.	Dr. Muhammad Munir	Medical Officer, Co-Opted Member, TAC, RIC, Rwp.
6.	Dr. Asad Ismail	Senior Medical Officer, Co-Opted Member, TAC, RIC, Rwp.
7.	Dr. Farhana Anwaar	Women Medical Officer, Co-Opted Member, TAC, RIC, Rwp.
8.	Eraj Islam	Pharmacist, Co-Opted Member, TAC, RIC, Rwp.
9.	Dr. Tauqeer Akbar	End User, Member, Associate Professor of Cardiac Surgery, RIC, Rwp.
10.	Dr. Musfireh Siddiqui	End User, Member, Associate Professor of Cardiac Surgery, RIC, Rwp.
11.	Dr. Qudsia Anjum Qureshi	End User, Member, Assistant Professor of Anesthesia, RIC, Rwp.
12.	Dr. Zainab Farid	End User, Member, Senior Registrar of Anesthesia, RIC, Rwp.
13.	Lt Col (R) Dr. Arif Maqsood Ali	End User, Member, Senior Microbiologist, RIC, Rwp.


Bids of following bidders were evaluated as per conditions/ criteria laid down in bidding documents of RIC. Items quoted by different companies were selected/rejected according to the End-Users report. After going through the specifications of items demanded in the bidding documents of RIC the End Users examined the provided items and satisfied with their registration, Authorization, Certification and Quality. The rejected items are not up to the specification/ quality. **Detailed comparative and Technical status of selected/rejected items along with the reasons of rejection by the End Users is mentioned in the Technical Comparative attached as Annex-A.**


*(Handwritten signatures and initials)*


S #	Name of Participating Firm	Total Marks	Qualifying Marks	Obtained Marks	Responsive/ Non-Responsive
1	M/s Medtronic Pakistan Pvt Ltd	65	65% (42.25/65)	53	Responsive for selected items.
2	M/s Iqbal & Company	65	65% (42.25/65)	53	Responsive for selected items.
3	M/s Amson Vaccines & Pharma	65	65% (42.25/65)	48	Responsive for selected items.
4	M/s Kohinoor Industries	65	65% (42.25/65)	53	Responsive for selected items.
5	M/s 3M Surgicals	65	65% (42.25/65)	33	Non-Responsive due to non provision of valid Drug sale license, past performance of the bidder & market experience of quoted items
6	M/s Freesia Enterprises	65	65% (42.25/65)	48	Responsive for selected items.
7	M/s Economic Health Enterprises	65	65% (42.25/65)	36	Non-Responsive due to non provision of past performance of the bidder, market availability of quoted items in leading chain stores & pharmacies for last two tears.
8	M/s Meher Traders	65	65% (42.25/65)	45	Responsive for selected items.
9	M/s Claris Medical	65	65% (42.25/65)	43	Responsive for selected items.
10	M/s Three 13 Pharma	65	65% (42.25/65)	53	Responsive for selected items.
11	M/s Kaumedex	65	65% (42.25/65)	49	Responsive for selected items.
12	M/s Nisa. SF (Pvt) Ltd	65	65% (42.25/65)	54	Non-Responsive due to rejection of quoted items.
13	M/s Nisa Impex (Pvt) Ltd	65	65% (42.25/65)	51	Responsive for selected items.
14	M/s 4A International	65	65% (42.25/65)	43	Responsive for selected items.
15	M/s Taha Enterprises	65	65% (42.25/65)	47	Non-Responsive due to rejection of quoted items.
16	M/s Farooq Agencies	65	65% (42.25/65)	45	Responsive for selected items.
17	M/s Al-Hamd Enterprises	65	65% (42.25/65)	48	Responsive for selected items.
18	M/s Sadqain Health Care Pvt Ltd	65	65% (42.25/65)	49	Responsive for selected items.
19	M/s Sanwa Trading Corporation	65	65% (42.25/65)	36	Non-Responsive due to non provision of documents on sr. 8 & 9 of compulsory parameters & market availability of quoted items in leading chain stores and pharmacies for last 02 years.
20	M/s Wah Nobel Acetates Ltd	65	65% (42.25/65)	43	Responsive for selected items.

21	M/s Mediland Pakistan (Pvt) Ltd	65	65% (42.25/65)	44	Responsive for selected items.
22	M/s Medicare Enterprises	65	65% (42.25/65)	43	Responsive for selected items.
23	M/s Verizon	65	65% (42.25/65)	49	Responsive for selected items.
24	M/s Cardiac Care	65	65% (42.25/65)	51	Responsive for selected items.
25	M/s FW Distributors	65	65% (42.25/65)	56	Responsive for selected items.
26	M/s UDL Distributors Pvt Ltd	65	65% (42.25/65)	46	Responsive for selected items.
27	M/s Flowtronix Systems	65	65% (42.25/65)	46	Responsive for selected items.
28	M/s Hakimson (Private) Limited	65	65% (42.25/65)	47	Responsive for selected items.
29	M/s FS Pharma	65	65% (42.25/65)	31	Non-Responsive due to non provision of Tax returns, annual turnover, list of technical staff & rejection of quoted items
30	M/s Moon Enterprises	65	65% (42.25/65)	43	Responsive for selected items.
31	M/s Anwar & Sons	65	65% (42.25/65)	44	Responsive for selected items.
32	M/s M Yousaf & Co	65	65% (42.25/65)	44	Responsive for selected items.
33	M/s Muller & Phipps (Pvt) Ltd	65	65% (42.25/65)	45	Responsive for selected items.
34	M/s Karim Industries	65	65% (42.25/65)	54	Responsive for selected items.
35	M/s Eastern Medical Care (Pvt) Ltd	65	65% (42.25/65)	51	Responsive for selected items.
36	M/s Cotton Craft (Pvt) Ltd	65	65% (42.25/65)	49	Responsive for selected items.
37	M/s Safeway Systems Pakistan	65	65% (42.25/65)	44	Responsive for selected items.
38	M/s Hope Pharma	65	65% (42.25/65)	48	Responsive for selected items.
39	M/s Usmanco International	65	65% (42.25/65)	56	Responsive for selected items.
40	M/s Next Pharmaceutical (Pvt) Ltd	65	65% (42.25/65)	54	Responsive for selected items.
41	M/s Magma	65	65% (42.25/65)	46	Responsive for selected items.
42	M/s Azam Trading Services	65	65% (42.25/65)	45	Responsive for selected items.
43	M/s Akram Brothers & co	65	65% (42.25/65)	46	Responsive for selected items.
44	M/s Surgimedika Pharma	65	65% (42.25/65)	14	Non-Responsive due to non provision of free sale certificate, sole agent, GMP, undertakings, ISO certification, list of technical staff, production capacity & undertaking of no batch failed during last 03 years
45	M/s A&E Medical	65	65% (42.25/65)	44	Responsive for selected items.


46	M/s KS Agencies	65	65% (42.25/65)	32	Non-Responsive due to non provision of market experience of quoted items, Tax returns & annual turnover
47	M/s Imran Traders	65	65% (42.25/65)	46	Responsive for selected items.
48	M/s Endoaid Biomedica	65	65% (42.25/65)	46	Substantial Responsive. Subject to provision of undertakings in compulsory parameters.
49	M/s Global Healthcare	65	65% (42.25/65)	46	Responsive for selected items.
50	M/S Popular International	65	65% (42.25/65)	51	Responsive for selected items.
51	M/S Brookes Pharma (Pvt) Ltd	65	65% (42.25/65)	52	Responsive for selected items.
52	M/S Ali & Sons	65	65% (42.25/65)	43	Responsive for selected items.


  
**Dr. Tazeer Akbar**  
 End User (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Dr. Musfirah Siddique**  
 End User (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi


  
**Dr. Qudsia Anjum Qureshi**  
 End User (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi


  
**Dr. Zahid Farid**  
 End User (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Lt Col (R) Dr. Arif Maqsood Ali**  
 End User (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Dr. Muhammad Munir**  
 Co-Opted (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi


  
**Dr. Farhana Anwaar**  
 Co-Opted (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Dr. Asad Ismail**  
 Co-Opted (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Eraj Islam**  
 Co-Opted (Member)  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**M. Noor-ul-Amin**  
 Member  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Dr. Humaira Farooq Kiani**  
 Member  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Dr. Gul e Raana Waseem**  
 Member/ Secretary  
 Technical Advisory Committee  
 RIC, Rawalpindi

  
**Dr. Javed Iqbal**  
 Chairman  
 Technical Advisory Committee  
 RIC, Rawalpindi


# RAWALPINDI INSTITUTE OF CARDIOLOGY

## Technical Advisory Committee

**Technical Comparative of Firms Participating for Annual Tender of Surgical & Disposable  
Items for OT, Anesthesia & Other Departments for FY 2022-23.**

**Ref No. RIC/PO/4466/22, Dated 24-03-2022**

Sr. No	Name of Item	Name of Firm	Registration Status	Manufacturer/ Brand	Selected/ Rejected
<b>OT &amp; PERFUSION DEPARTMENTS</b>					
<b>MONOFILAMENT POLYPROPYLENE</b>					
	Monofilament polypropylene 4/0 round body double needle	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001436 2020	Johnson & Johnson, Ethicon	Selected
		M/S Anwar & Sons	MDIR-0002870 2021	Polypropylene, SMI, Belgium	Selected
		M/S Akram Brothers & Co	MDIR-0003098 2022	Wego Prolene 4/0, 16mm, 17mm, 1/2 Circle Double CC Round Bodied Needle, Fousin Medical Supplies Inc. Ltd. China	Rejected
		M/S Popular International	MDIR-0000858 2019	Surgipro, Covidien Medtronic USA	Selected
		M/S Magma	MDMR-000045 2018	Surgiline, Vikor Healthcare Pvt Ltd, Pakistan	Selected
2	Monofilament polypropylene 5/0 round body double needle	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001436 2020	Johnson & Johnson, Ethicon	Selected
		M/S Cardiac Care	0003164 18-01-2022	Peter, France	Selected
		M/S Anwar & Sons	MDIR-0002870 2021	Polypropylene, SMI, Belgium	Selected
		M/S Akram Brothers & Co	MDIR-0003098 2022	Wego Prolene 5/0, 16mm, 17mm, 1/2 Circle Double Round Bodied, Fousin Medical Supplies Inc. Ltd. China	Rejected
		M/S Popular International	MDIR-0000858 2019	Surgipro, Covidien Medtronic USA	Selected
3	Monofilament polypropylene 5/0 round body double needle	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001436 2020	Johnson & Johnson, Ethicon	Selected
		M/S Cardiac Care	0003164 18-01-2022	Peter, France	Selected
		M/S Anwar & Sons	MDIR-0002870 2021	Polypropylene, SMI, Belgium	Selected
		M/S Akram Brothers & Co	MDIR-0003098 2022	Wego Prolene 5/0, 13mm, 1/2 Circle Double Round Bodied, Fousin Medical Supplies Inc Ltd China	Rejected
		M/S Popular International	MDIR-0000858 2019	Surgipro, Covidien Medtronic USA	Selected



4	Monofilament polypropylene 5/0 round body double needle	M/S Akram Brothers & Co	MDIR-0003098 2022	Wega Proline 5/0, 11mm, 1/2 Circle Double Round Hooked needle, Fosun Medical Supplies Inc. Ltd China	Rejected
		M/S Popular International	MDIR-0000858 2019	Surgipro, Covidien Medtronic USA	Selected
5	Monofilament polypropylene 6/0 round body double needle	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001436 2020	Johnson & Johnson, Ethicon	Selected
		M/S Anwar & Sons	MDIR-0002870 2021	Polypropylene, SMI, Belgium	Rejected Needle bigger
		M/S Akram Brothers & Co	MDIR-0003098 2022	Wega Proline 6/0, 8mm, 6/0, 9mm, 11mm 1/2 Circle Round Hooked Double, Fosun Medical Supplies Inc. Ltd China	Rejected
		M/S Popular International	MDIR-0000858 2019	Surgipro, Covidien Medtronic USA	Selected
6	Monofilament Prolene 2-0, straight needle, rounded	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001436 2020	Johnson & Johnson, Ethicon	Selected
		M/S Cardiac Care	0003164 18-01-2022	Peter, France	Selected
		M/S Akram Brothers & Co	MDIR-0003098 2022	Wega Proline Polypropylene 2-0, 60mm, Straight Cutting, Fosun Medical Supplies Inc. Ltd China	Selected
		M/S Popular International	MDIR-0000858 2019	Surgipro, Covidien Medtronic USA	Selected
		M/S Magma	MDMR-000045 2018	Surgiline, Vikor Healthcare Pvt Ltd, Pakistan	Selected
		<b>POLYSTER</b>			
7	Polyster 2/0 round body double needle with pledget (Valve Suture)	M/S Cardiac Care	0003165 18-01-2022	Peter, France	Selected
		M/S Popular International	MDIR-0000849 2017	Ticon, Covidien Medtronic USA	Selected
8	Polyster 2/0 round body double needle without pledget (Free)	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001432 2020	Johnson & Johnson, Ethicon	Selected
		M/S Popular International	MDIR-0000849 2017	Ticon, Covidien Medtronic USA	Selected
9	Polyster 2/0 round body double needle without pledget (Free)	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001432 2020	Johnson & Johnson, Ethicon	Selected
		M/S Popular International	MDIR-0000849 2017	Ticon, Covidien Medtronic USA	Selected
10	Polyster 2/0 round body double needle with pledget (Valve Suture)	M/S Cardiac Care	0003165 18-01-2022	Peter, France	Selected
		M/S Popular International	MDIR-0000849 2017	Ticon, Covidien Medtronic USA	Selected

POLYGLACTIN					
11	Braided absorbable polyglactin suture 3/0 cutting Needle	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001430 2020	Johnson & Johnson, Ethicon	Selected
		M/S Akram Brothers & Co	MDIR-0003207-A 2022	Wego PVA Poly glycolic Acid/PGA Poly (Lactic-co-glycolic acid) 9/0 size 3/0, 16mm, cutting needle, Fousun Medical Supplies Inc. Ltd China	Selected
OT ITEMS					
12	Bone wax	M/S Akram Brothers & Co	C0830 2022	Bone Waxx2.5g, Brasuture	Selected
		M/S Popular International	MDIR-0001360 2020	Bonewax, Covidien Medtronic USA	Selected
13	Presterilized Cardiac sponges	M/S Cotton Craft (Pvt) Ltd	MDMR-001971	Local Manufacturer	Selected by affidavit provided register
14	Chest tube	M/S Usman Co International	MDIR-0003044	Thoracic Catheter Without Trocar, Bicakcilar Dis Ticaret A S, Turkey	Selected
		M/S Cardiac Care	Not Registered	Pahsco, Taiwan	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected Too hard
15	Diathermy pencil	M/S Meher Traders	Not Applied	Wellmed, Zhejiang Jinhua Huatong China	Rejected Poor quality switch stiff
		M/S Cardiac Care	Not Registered	Fiab, Italy	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Selected
		M/S M. Yousaf & Co	Applied 30-04-2021	MYKU Disposable Electrosurgical Control Pencil, Bismed International Pakistan	Rejected Stiff switch
16	Scratch pad	M/S Flowtronix Systems	Not Applied	TPRI, China	Selected
		M/S Cardiac Care	Not Registered	TPRI, China	Selected
17	Nylon tapes				
18	Y connector / Connector (all sizes with leur lock)	M/S Flowtronix Systems	Not Applied	TPRI, China	Selected
		M/S Cardiac Care	MDIR-0003163 18-01-2022	Euroset, Italy	Selected



13

19	Yanker set	M/S Freesia Enterprises	Applied 2019	Canack	Selected
		M/S Meher Traders	Applied 02-12-2020	Wellmed, Hingbo Guangzhou Medical China	Rejected Hole on tip
		M/S Usman Co International	MDIR-0003047	Yankner Yankner Connecting Set - Handle With Top Cut In For a Vacuum, Hicakolar Enk Tuzet A S Turkey	Rejected
		M/S Cardiac Care	Applied Nov-2021	Greenmed, China	Selected
		M/S Hakimsons (Pvt) Ltd		Freyemed, Hingbo MFLAH Medical China	Rejected Safe hole not required
		M/S M. Yousaf & Co	Applied 30-04-2021	Skyman Cardiac Surgery Disposable Cannula, Skybro Health Care Pakistan	Selected
20	Level sensor holder	M/S Flowtronix Systems	Not Applied	TPRI, China	Rejected Not compatible
		M/S Cardiac Care	Not Applied	TPRI, China	Selected

**PERFUSION DISPOSABLES ITEMS FOR ADULT & PEADS**

**ADULT DISPOSABLES**

21	Aortic cannula straight tip	M/S Flowtronix Systems	MDIR-0001521 09-06-2020	Andacor, Belgium	Selected
		M/S Safeway Systems Pakistan	MDIR-0001779 16-09-2020	Cardiomed Supplies Inc, Canada	Rejected
		M/S Medtronic Pakistan	MDIR-0001005 2019	DLP Flexible Arch Cannula, Medtronic Minneapolis, MN USA	Selected
22	Femoral arterial cannula	M/S Cardiac Care	Applied March-2022	EDWARD Life, USA	Rejected SNP
		M/S Medtronic Pakistan	MDIR-0001322 2020	Biomedicus One Piece Femoral Arterial Cannulae, Minneapolis, MN, USA	Selected
23	Femoral venous cannula multi stage	M/S Cardiac Care	Applied March-2022	EDWARD Life, USA	Rejected SNP
		M/S Medtronic Pakistan	MDIR-0001324 2020	Biomedicus One Piece femoral Arterial Cannulae, Minneapolis, MN, USA	Selected
24	Aortic cannula angle tip	M/S Flowtronix Systems	MDIR-0001521 09-06-2020	Andacor, Belgium	Selected
		M/S Safeway Systems Pakistan	MDIR-0001779 16-09-2020	Cardiomed Supplies Inc, Canada	Rejected
		M/S Medtronic Pakistan	MDIR-0000172 2019	DLP Arterial Curved Cannula, Medtronic Minneapolis, MN USA	Selected
25	Ascending aortic valsalva valve conduits	M/S Verizon	MDIR-0000234 2019	Ascending Aortic Prothesis, St. Jude Medical, USA	Selected
		M/S Cardiac Care	0001340 08-02-2020	On-X Cryolife, USA	Selected
26	Biogluce 5ml or equivalent	M/S Cardiac Care	MDIR-0002075 25-01-2021	Cryolife, USA	Selected
27	Bovine pericardial patch	M/S Verizon	MDIR-0002774 2021	SIM Percutial Patch, St. Jude Medical, USA	Selected
		M/S 3M Surgical	Not Given	Bovv Patch, HP Dho Ltda Brazil	Selected

Handwritten signatures and initials are present at the bottom of the page, including a circled number '14'.



28	Cardiac sump	M/S Flowtronix Systems	MDIR-0001069 13-12-2019	Andacor, Belgium	Selected
		M/S Medtronic Pakistan	MDIR-0000174,156 2019	DLP, Medtronic, USA	Selected
29	Contegra conduit graft / double valvor	M/S Medtronic Pakistan	MDIR-0000158 2019	Contegra, Medtronic, USA	Selected
30	Coronary ostial cannula	M/S Flowtronix Systems	MDIR-0001070 13-12-2019	Andacor, Belgium	Selected
		M/S Medtronic Pakistan	MDIR-0001004 2019	DLP Coronary Artery ostial cannula, Medtronic Minneapolis, MN USA	Selected
31	Coronary shunts	M/S Medtronic Pakistan	MDIR-0001334 2020	Clearview Intracoronary Shunt, Medtronic USA/Mexico	Selected
32	Dacron patch sheet	M/S 3M Surgical	Not Given	Knitted Vascular Patch with collagen, VUP Medical CZ	Rejected Small size
		M/S Cardiac Care	MDIR-0001303 13-02-2020	BARD, USA	Selected
33	Diathermy disposable pads with leads	M/S Cardiac Care	Not Applied	Fiab, Italy	Rejected No Lead
		M/S A & E Medical	Not Applied	Grounding Pads CE Mark, Fiab-Spa, Italy	Selected
		M/S Hakimsons (Pvt) Ltd		Swaroplate, Nissba Medical Technologies Ltd, UK	Rejected Lead not compatible
34	Fogarty catheter	M/S Cardiac Care	Applied Nov-2020	Perouse Medical (Vsgon) Europe	Selected
		M/S Imran Traders	Applied	Hagmad, Hagmad Poland	Selected
35	Hemofiltration set	M/S Flowtronix Systems	MDIR-0002886 18-10-2021	Brizio, Nipro, Japan	Selected
		M/S Cardiac Care	MDIR-000998 06-12-2019	Medica, Italy	Selected
36	Intra aortic balloon	M/S UDL Distribution (Pvt) Ltd	MDIR-0000707 16-09-2019	Arrow, USA (Teleflex)	Rejected Hole is too long, bleeds early
		M/S Cardiac Care	MDIR-0000738 27-19-2019	Insightra, USA	Selected
37	LV vent	M/S Flowtronix Systems	MDIR-0001522 09-06-2020	Andacor, Belgium	Selected
		M/S Medtronic Pakistan	MDIR-0001008 2019	DLP left heart vent catheter, Medtronic Minneapolis, MN USA	Selected
38	PTFE felt sheet	M/S 3M Surgical	Not Given	Cortex I, VUP Medical CZ	Rejected Poor Quality
		M/S Cardiac Care	MDIR-0001304 13-02-2020	BARD, USA	Selected
39	PTFE sutures				
40	Aortic Valve bileaflet reduced ring for (small annulus)	M/S Verizon	MDIR-0000235 2019	HP Rotatable Heart valve aortic, St. Jude Medical, USA	Selected
		M/S Cardiac Care	MDIR-0000458 24-06-2019	On-X Cryolife, USA	Selected
		M/S Medtronic Pakistan	MDIR-0001017 2019	Medtronic Open Pivot (AP aortic) Minneapolis, MN USA	Selected

Handwritten signatures and initials are present at the bottom of the page, including a circled number '15'.

41	Retrograde cardioplegia cannula with wire	M/S Flowtronix Systems	MDIR-0001528 09-06-2020	Andacor, Belgium	Selected
		M/S Cardiac Care	MDIR-0003318 16-03-2022	EDWARD Life, USA	Selected
		M/S Medtronic Pakistan	MDIR-0000578 2019	DLP PVC RCSP Cannulae with Auto inflate Cuff, Medtronic Minneapolis, MN USA	Selected
42	Sternum saw blade with code/machine	M/S Claris Medical	Not Applied	De-Soutter Medical, UK	Selected
43	Annuloplasty tricuspid ring (Flexible / Semi Flexible)	M/S Verizon	MDIR-0000241 2019	Flexible Valve ring Tricuspid, St. Jude Medical, USA	Selected
		M/S Medtronic Pakistan	MDIR-0000163 2019	Duran Ancore Annuloplasty ring, Medtronic Santa, CA USA	Selected
44	Annuloplasty mitral ring (Flexible / Semi Flexible)	M/S Verizon	MDIR-0000241 2019	Flexible Valve ring Mitral, St. Jude Medical, USA	Selected
		M/S Medtronic Pakistan	MDIR-0000161 2019	CG Future Annuloplasty ring, Medtronic Santa, CA USA	Selected
45	Mitral ring 3D	M/S Medtronic Pakistan	MDIR-0000815 2019	Profile 3D Annuloplasty ring, Medtronic USA/Mexico	Selected
46	Venous saturation connector	M/S Flowtronix Systems	MDIR-0001519 09-06-2020	1/2*1/2, TPRI, China	Selected
47	Venous saturation connector	M/S Flowtronix Systems	MDIR-0001519 09-06-2020	3/8*3/8, TPRI, China	Selected
48	Heart stabilizer octopus	M/S Medtronic Pakistan	MDIR-0000167 2019	Octopus Evolution Tissue Stabilizer, Medtronic Minneapolis, MN USA	Selected
49	Surgical absorbable hemostat	M/S Muller & Phipps (Pvt) Ltd	MDIR-0001443 2020	(5x35cm, 10x20cm) Johnson & Johnson, Ethicon	Selected
		M/S K.S Agencies	Applied Nov-2020	Oxigel Standard 5x35cm / 10x20cm, Beta Tech Turkey	Selected
50	Urchin / star fish	M/S Medtronic Pakistan	MDIR-0002256, 1329 2021	Starfish Evo Heart Positioner/Urchin Evo Heart Positioner, Medtronic Minneapolis, MN USA	Selected
51	Dacron tube graft	M/S 3M Surgical	Not Given	Incompressible Knitted vascular prosthesis with collagen, VUP Medical CZ	Selected
		M/S Cardiac Care	MDIR-000733 27-09-2019	Atrium/Getting, USA	Rejected SNP
52	Redivas drains	M/S Imran Traders	Not Applied	Privac, Primed Germany	Selected
53	Cardioplegia delivery System with coil adult and peads	M/S Flowtronix Systems	MDIR-0001068 13-12-2019	TPRI, China	Selected
		M/S Cardiac Care	MDIR-0002855 14-09-2021	Euroset, Italy	Selected With modified for circuit compatibility

**PAEDS DISPOSABLES**

54	Aortic cannula straight tip	M/S Safeway Systems Pakistan	MDIR-0001779 16-09-2020	Cardiomed Supplies Inc, Canada	Selected with cap
		M/S Medtronic Pakistan	MDIR-000154 2019	DLP Straight Tip Atrial Cannula, Medtronic Minneapolis, MN USA	Selected
55	Aortic cannula paed's angle tip	M/S Safeway Systems Pakistan	MDIR-0001779 16-09-2020	Cardiomed Supplies Inc, Canada	Selected with cap
56	Metal tip right angle venous cannula	M/S Flowtronix Systems	MDIR-0001519 09-06-2020	Andacor, Belgium	Selected
		M/S Safeway Systems Pakistan	MDIR-0001780 16-09-2020	Cardiomed Supplies Inc, Canada	Selected
		M/S Medtronic Pakistan	MDIR-0002273, 1007 2019	DLP Single stage Metal Tip Venous Cannula, Medtronic Minneapolis, MN USA	Selected
57	Single stage renal cannula wire reinforced bullet tip French	M/S Flowtronix Systems	MDIR-0001519 09-06-2020	Andacor, Belgium	Selected
		M/S Medtronic Pakistan	MDIR-0001006 2019	DLP Single Stage Venous Cannulae, Medtronic Minneapolis, MN USA	Selected
58	Femoral arterial cannula	M/S Medtronic Pakistan	MDIR-0001323, 1322 2020	Biomedicus One Piece femoral Arterial Cannulae, Minneapolis, MN, USA	Selected
59	Femoral venous cannula French	M/S Medtronic Pakistan	MDIR-0001931, 1324 2020	Biomedicus One Piece femoral Arterial Cannulae, Minneapolis, MN, USA	Selected
60	Antegrade cardioplegia	M/S Flowtronix Systems	MDIR-0001517 09-06-2020	Andacor, Belgium	Selected
		M/S Medtronic Pakistan	MDIR-0001010, 1022 2019	DLP Left Heart vent catheter/DLP Aortic Root Cannula, Medtronic Minneapolis, MN USA	Selected
61	Hemofiltration set	M/S Flowtronix Systems	MDIR-0002886 18-10-2021	Brizio, Nipro, Japan	Selected
		M/S Cardiac Care	MDIR-0000998 06-12-2019	Medica, Italy	Selected
62	PTFE tube graft	M/S 3M Surgical	Not Given	Knitted Vascular prosthesis with collagen, VUP Medical CZ	Rejected SNP
		M/S Cardiac Care	MDIR-0000732 27-09-2019	Atrium/Getinge, USA	Rejected SNP
63	Epicardial lead for PPM with generator				

**MISCELLANEOUS DISPOSABLES**

64	EVH Kits	M/S Mediland Pakistan	Applied 2018	Vasoview Hemopro 2 Harvesting kit (VH-4000), Getinge Maquet, USA	Selected
65	ECMO kits cardio help	M/S Mediland Pakistan	Applied 2018	BE HLS-7050, Getinge Maquet, Germany	Selected
66	ECMO rota flow	M/S Mediland Pakistan	Applied 2018	BE-PLS 2050, Getinge Group, Germany	Selected
67	Transit time flow meter	M/S Claris Medical	Not Applied	Medistim TTFM Probe 1.5mm, 2.0mm, 4.0mm with handle, Medistim ASA, Norway	Selected

Handwritten signatures and initials are present below the table, including a circled number '17'.

ANESTHESIA & ITC DEPARTMENT					
68	CVP line Quad Lumen (Adult)	M/S UDL Distribution (Pvt) Ltd	MDIR-0000437 19-06-2019	Arrow, USA (Teleflex)	Selected
		M/S Iqbal & Company	MDIR-0000320 2019	Medcomp, USA	Selected
		M/S Flowtronix Systems	MDIR-0001067 13-12-2019	ABLE, Guangdong Baibe Medical, China	Rejected
		M/S Cardiac Care	MDIR-0000737 27-09-2019	Vygon, Europe	Selected
69	CVP line Tripple Lumen (Peads)	M/S UDL Distribution (Pvt) Ltd	MDIR-0000437 19-06-2019	Arrow, USA (Teleflex)	Selected
		M/S Iqbal & Company	MDIR-0000321 2019	Medcomp, USA	Selected
		M/S Flowtronix Systems	MDIR-0001067 13-12-2019	ABLE, Guangdong Baibe Medical, China	Rejected
		M/S Cardiac Care	MDIR-0000736 27-09-2019	Vygon, Europe	Selected
70	Double DPT Kit	M/S FW Distributor	Applied 18-06-2021	YH Medical	Selected
		M/S F.S Pharma	Applied	SCW China Sole authorized distributor certificate	Rejected SNP
		M/S Flowtronix Systems	MDIR-0001076 13-12-2019	ABLE, Guangdong Baibe Medical, China	Rejected Telephonic connector not compatible
		M/S Safeway Systems Pakistan	MDIR-0002842 10-09-2021	Bicukular Global Turkey	Rejected Not compatible
		M/S Cardiac Care	Applied 09-06-2021	SCW Medical China Sole authorized distributor certificate	Selected
71	Arterial Cath 18G	M/S Cardiac Care	MDIR-0000596 23-07-2019	Vygon, Europe	Selected
72	Bougies (Adult)	M/S Economic Health Enterprises	Not Applied	Smith's Medical International	Rejected SNP
		M/S Cardiac Care	MDIE-0000220 15-02-2021	Vygon, Europe	Selected
73	Bougies (Peads)	M/S Cardiac Care	MDIE-0000220 15-02-2021	Vygon, Europe	Selected
74	Laryngeal mask airways (LMA) disposable	M/S 4A International	MDIR-0002295 2021	Emedical Laryngeal Mask Airway, Excellence Medical, China	Selected
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0000948	Sikas, Neomate, Laryngeal mask airway, Intersurgical	Selected
		M/S Hakinsons (Pvt) Ltd	MDIR-0002809 30-08-2021	Frymed, Nangbo MFLAB Medical, China	Selected
<b>ITC DISPOSABLE ITEMS</b>					
75	Warmer blanket	M/S Economic Health Enterprises	Not Applied	Smith's Medical International	Rejected
		M/S Cardiac Care	Not Applied	Celsisu, Spain	Selected
		M/S Magma	Applied	Patient Warming Blanket 30000, 3M, USA	Selected

Handwritten signatures and initials are present at the bottom of the page, including a circled number 18.

76	DPT kit single	M/S FW Distributor	Applied 18-06-2021	YH Medical	Selected
		M/S F.S Pharma	Applied	SCW China	Rejected SNP
		M/S Flowtronix Systems	MDIR-0001076 13-12-2019	ABLE, Guangdong Baibe Medical, China	Rejected
		M/S Safeway Systems Pakistan	MDIR-0002842 10-09-2021	Bicakcilar Global Turkey	Rejected Not compatible
		M/S Cardiac Care	Applied 09-06-2021	SCW Medicath, China	Selected
77	CPAP mask (Adult)	M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0000950 2019	Scented, anaesthetic face mask with green hook ring, Intersurgical	Selected
		M/S Cardiac Care	MDIR-0002630 29-06-2021	Vygon, Europe	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected
78	CPAP mask (Peds)	M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0000950 2019	Scented, anaesthetic face mask with green hook ring, Intersurgical	Selected
		M/S Cardiac Care	MDIR-0002630 29-06-2021	Vygon, Europe	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected SNP
79	CPAP mask (Neonate)	M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0000950 2019	Scented, anaesthetic face mask with green hook ring, Intersurgical	Selected
		M/S Cardiac Care	MDIR-0002630 29-06-2021	Vygon, Europe	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected SNP
80	O2 Mask Disposable (Adult)	M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001161	Adult, Medium concentration oxygen mask with nose clip, Intersurgical	Selected
		M/S Meher Traders	MDIR-0002371 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	MDIR-0001788 16-09-2020	Safety	Selected
		M/S Cardiac Care	Applied 07-01-2020	Hsincr, Taiwan	Rejected Connector too hard
		M/S Hakimsons (Pvt) Ltd	MDIR-0002808 30-08-2021	Foyomed, Ningbo MFLAB Medical/ China	Selected
		M/S Azam Trading Services	Applied 26-03-2020	Ultramed Oxygen Mask Adult, Ultra for Medical, Egypt	Selected
81	O2 Mask Disposable (Peds)	M/S Meher Traders	MDIR-0002371 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	MDIR-0001788 16-09-2020	Safety	Rejected SNP
		M/S Cardiac Care	Applied 07-01-2020	Hsincr, Taiwan	Rejected SNP
		M/S Hakimsons (Pvt) Ltd	MDIR-0002808 30-08-2021	Foyomed, Ningbo MFLAB Medical/ China	Selected
		M/S Azam Trading Services	Applied 26-03-2020	Ultramed Oxygen Mask Peds, Ultra for Medical, Egypt	Selected

82	O2 Mask Disposable (Neonate)	M/S Farooq Agencies	MDIR-0001788 16-09-2020	Safety	Rejected SNP
83	Rebreathing mask (Adult)	M/S 4A International	MDIR-0002297 2021	EM01-006, Excellence Medical, China	Rejected Nozzle too hard
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001161	Adult, high concentration oxygen mask with tube, Intra-surgical	Selected
		M/S Meher Traders	Applied 07-06-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	Not Applied	Safety	Selected
		M/S Cardiac Care	Not Applied	Haitur, Taiwan	Rejected SNP
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Selected
84	Rebreathing mask (Peds)	M/S 4A International	MDIR-0002297 2021	EM01-007, Excellence Medical, China	Rejected Nozzle too hard to attached with flowmeter
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001315 2020	Paediatric, high concentration oxygen mask with tube, Intra-surgical	Selected
		M/S Meher Traders	Applied 07-06-2021	Wellmed, Ningbo Greatmountain Medical China	Rejected SNP
		M/S Farooq Agencies	Not Applied	Safety	Rejected SNP
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected Connector faulty
85	Nasal cannula / Nasal prong	M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001153 2020	Adult, nasal cannula with straight prongs and tube, Intra-surgical	Selected
		M/S Meher Traders	MDIR-0002373 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	MDIR-0001790 16-09-2020	Safety	Selected
		M/S Hakimsons (Pvt) Ltd	MDIR-0003325 15-04-2022	Foyomed, Ningbo MFLAB Medical/ China	Selected
86	PD catheter				
87	Butter fly needle				
88	Chest binder	M/S Three13 Pharma	Not Applied	Local	Rejected SNP
		M/S Flowtronix Systems	Not Applied	Local, Pakistan	Rejected
		M/S Cardiac Care	Not Applied	Local	Selected
		M/S Imran Traders	Not Applied	Perfect, Qadri Healthco Pakistan	Selected
89	Round bottle single	M/S Cardiac Care	Not Applied	Pahsco, Taiwan	Selected
90	Opsite Dressing	M/S K.S Agencies	Not Applied	Hi-Tec Surgical Incise Drape 15*28cm, 45*55cm, Klas Medikal Turkey	Selected
		M/S Cardiac Care	Applied April-2021	Vygon, Europe	Selected
		M/S Magma	MDIR-0001080 2019	loban 8840, JM, USA	Selected

91	Duoderm patch	M/S Cardiac Care	Not Applied	T.Lab Turkey	Selected
92	Dial a flow	M/S Three 13 Pharma	MDIR-0002203 2021	KAWA	Selected
		M/S Surwa Trading Corporation	Not Given	China	Selected
		M/S Meher Traders	Applied 06-05-2021	Wellmed, Suzhou Health Plastic Products China	Selected
		M/S Farooq Agencies	Applied 13-09-2019	Safety	Selected
		M/S Usman Co International	MDIR-0000162	B-Flow, Bicaclar Di Ticaret A.S, Turkey	Rejected
		M/S Cardiac Care	MDIR-0001646 12-08-2020	Pro-Active, Italy	Selected
		M/S Hakimsons (Pvt) Ltd		Fuyomed, Ningbo MFLAB Medical/ China	Selected
93	External catheter				
94	Cora pore dressing Or equivalent	M/S Three 13 Pharma	MDIR-0000178 2021	OXYMAX	Selected
		M/S Cardiac Care	Not Applied	China	Selected
		M/S Inran Traders		Protectpore China	Selected
95	Fluid warmer set	M/S Cardiac Care	Not Applied	Biegler, Austria	Selected
		M/S Magma	Applied	Ranger Fluid Warming Set 24250, 3M, USA	Selected
96	CRRT Kit compatible with / Fluid RIC CRRT machine				
97	CRRT disposables				
<b>C OTHERS DEPARTMENT / WARDS</b>					
98	ETT Adult Cuffed	M/S Freesia Enterprises	Applied 2019	Canack	Selected
		M/S Meher Traders	MDIR-0002367 29-03-2021	Wellmed, Nanchang Kamed Medical China	Rejected No previous experience
		M/S Farooq Agencies	Applied 07-09-2020	PIGEON	Rejected Connection very loose Early dislodges
		M/S Usman Co International	MDIR-0002429	Endosoft Endotracheal Tube Cuff, Cholecom Medical Devices Co. Ltd, China	Selected
		M/S Hakimsons (Pvt) Ltd	Applied 01-06-2020	Fuyomed, Ningbo MFLAB Medical/ China	Selected
99	ETT Peds Cuffed	M/S Freesia Enterprises	Applied 2019	Canack	Selected
		M/S Meher Traders	MDIR-0002367 29-03-2021	Wellmed, Nanchang Kamed Medical China	Rejected Poor quality
		M/S Usman Co International	MDIR-0002429	Endosoft Endotracheal Tube Cuff, Cholecom Medical Devices Co. Ltd, China	Selected
		M/S Hakimsons (Pvt) Ltd	Applied 01-06-2020	Fuyomed, Ningbo MFLAB Medical/ China	Selected

100	ETT Peads Uncuffed	M/S Freesia Enterprises	Applied 2019	Canack	Selected
		M/S Meher Traders	MDIR-0002367 29-03-2021	Wellmed, Nanchang Kaiimed Medical China	Rejected Poor quality
		M/S Farooq Agencies	Applied 07-09-2020	PIGEON	Rejected Airway compromised tip is wider too soft to intubate
		M/S Usman Co International	Applied	Endosoft Endotracheal Tube Uncuff, Chiecom Medical Devices Co Ltd, China	Selected
		M/S Hakimsons (Pvt) Ltd	Applied 01-06-2020	Foyomed, Ningbo MFLAB Medical China	Selected
101	Intubation stylet (Adult)	M/S Cardiac Care	Not Applied	Pahsco, Taiwan	Rejected SNP
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical China	Selected
102	Intubation stylet (Peads)	M/S Cardiac Care	Not Applied	Pahsco, Taiwan	Rejected Poor quality too hard
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical China	Selected
103	HME Filter (Adult)	M/S 4A International	MDIR-0002309 2021	EM12-201, Excellentcare Medical, China	Selected
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0000286 2019	Inter-Therm HMEF with luer port-sterile, Intersurgical	Selected
		M/S Freesia Enterprises	Applied 2019	Canack	Selected
		M/S F.S Pharma	MDIR-0001411 19-08-2020	RVENT Turkey	Rejected No prior experience
		M/S Meher Traders	MDIR-0002374 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Rejected Poor quality
		M/S Cardiac Care	Applied 15-10-2019	Hsiner, Taiwan	Rejected SNP
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical China	Rejected Quality issue
104	HME Filter (Peads)	M/S 4A International	MDIR-0002309 2021	EM12-206, Excellentcare Medical, China	Selected
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001320 2020	Clear-Therm Mini HMEF with Luer port, Intersurgical	Selected
		M/S F.S Pharma	MDIR-0001411 19-08-2020	RVENT Turkey	Rejected No prior experience
		M/S Meher Traders	MDIR-0002374 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Rejected Poor quality
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical China	Rejected Quality issue
105	IV Cannula with wings 14G	M/S Farooq Agencies	074668 20-05-2019	JRZ LEBANON	Selected
		M/S AL-Hamd Enterprises	MDIR-0002264	Inpharves, Farconake for Advanced Medical Industries (S.A.E) Egypt	Selected
		M/S Usman Co International	MDIR-0001763	B-Cat 2 I/V Cannula, Bicalcular Da Tarant A 5 Turkey	Selected
		M/S Azam Trading Services	074713 13-03-2020	Q Flow I/V Cannula-wing type with injection port, Qutim German for medical Devices, Qatar	Selected



116	IV Cannula with wings 18G	M/S Farooq Agencies	17468 20/05-2021	202133281426	Selected
		M/S Al-Hamd Enterprises	MIDIR-1112264	Inpharven, Farcomake for Advanced Medical Industries (S.A.E.) Egypt	Selected
		M/S Usman Co International	MIDIR-1111763	IS-Cat 21 V Cannula Bucalular Ltd. Turkey A.S Turkey	Selected
		M/S Azam Trading Services	174713 13/03-2021	IS-Cat 21 V Cannula type with injection port. Farcomake for medical services, Egypt	Selected
117	IV Cannula with wings 18G	M/S Thero13 Pharma	MIDIR-111157 2021	VEINOFIX	Rejected Poor quality
		M/S Nisa Impex (Pvt) Ltd	MIDIR-1111277	Nisa Impex	Selected
		M/S Global Health Care	176189 2020	Ter Venocath Plus, Triage Medical Germany	Rejected Too much difficult for passing IV cannula due to soft tip
		M/S Farooq Agencies	17468 20/05-2021	202133281426	Selected
		M/S Al-Hamd Enterprises	MIDIR-1112264	Inpharven, Farcomake for Advanced Medical Industries (S.A.E.) Egypt	Selected
		M/S Usman Co International	MIDIR-1111763	IS-Cat 21 V Cannula, Bucalular Ltd. Turkey A.S Turkey	Selected
118	IV Cannula with wings 20G	M/S Imran Traders	MIDIR-111127 15/01-2021	Green, Silver Surgical Complex Pakistan	Rejected SNP
		M/S Azam Trading Services	MIDIR-1112277 13/03-2021	Vasofix 20G, B Braun Pakistan Pvt Ltd	Selected
		M/S Thero13 Pharma	MIDIR-111157 2021	VEINOFIX	Rejected Poor quality
		M/S Nisa Impex (Pvt) Ltd	MIDIR-1111277	Nisa Impex	Selected
		M/S Global Health Care	176189 2020	Ter Venocath Plus, Triage Medical Germany	Rejected Too much difficult for passing IV cannula due to soft tip
		M/S Farooq Agencies	17468 20/05-2021	IRZ, LEHANNON	Selected
		M/S Al-Hamd Enterprises	MIDIR-1112264	Inpharven, Farcomake for Advanced Medical Industries (S.A.E.) Egypt	Selected
M/S Usman Co International	MIDIR-1111763	IS-Cat 21 V Cannula Bucalular Ltd. Turkey A.S Turkey	Selected		
M/S Imran Traders	MIDIR-111127 15/01-2021	Green, Silver Surgical Complex Pakistan	Rejected Poor quality		
M/S Azam Trading Services	MIDIR-1112277 13/03-2021	Vasofix 20G, B Braun Pakistan Pvt Ltd	Selected		

109	IV Cannula with wings 22G	M/S Three13 Pharma	MDIR-000157 2021	VEINOFIX	Rejected Poor quality
		M/S Nisa Impex (Pvt) Ltd	MDIR-0001277	Nisa Impex	Selected
		M/S Global Health Care	066189 2020	Tor Venocath Plus, Troge Medical Germany	Rejected Too much difficult for passing I V cannula due to soft tip
		M/S Farooq Agencies	074668 20-05-2019	JRZ LEBANON	Selected
		M/S AL-Hamd Enterprises	MDIR-0002264	Inpharven, Farcomake for Advanced Medical Industries (S A E) Egypt	Selected
		M/S Usman Co International	MDIR-0001763	B-Cat 2 I V Cannula, Bicakcilar Dis Ticaret A.S Turkey	Selected
		M/S Imran Traders	MDMR-000127 05-07-2021	Green, Silver Surgical Complex Pakistan	Rejected Poor quality
		M/S Azam Trading Services	MDIR-0002277 10-03-2021	Vasofix 22G, B Braun Pakistan Pvt Ltd	Selected
110	IV Cannula with wings 24G	M/S Three13 Pharma	MDIR-000157 2021	VEINOFIX	Rejected Poor quality
		M/S Nisa Impex (Pvt) Ltd	MDIR-0001277	Nisa Impex	Selected
		M/S Global Health Care	066189 2020	Tor Venocath Plus, Troge Medical Germany	Rejected Too much difficult for passing I V cannula due to soft tip
		M/S Farooq Agencies	074668 20-05-2019	JRZ LEBANON	Selected
		M/S AL-Hamd Enterprises	MDIR-0002264	Inpharven, Farcomake for Advanced Medical Industries (S.A E) Egypt	Selected
		M/S Usman Co International	MDIR-0001763	B-Cat 2 I V Cannula, Bicakcilar Dis Ticaret A.S Turkey	Selected
		M/S Imran Traders	MDMR-000127 05-07-2021	Green, Silver Surgical Complex Pakistan	Rejected Poor quality
		M/S Azam Trading Services	MDIR-0002277 10-03-2021	Vasofix W 24G, B Braun Pakistan Pvt Ltd	Selected
111	IV cannula without wings	M/S Imran Traders	MDMR-000127 05-07-2021	Green, Silver Surgical Complex Pakistan	Selected
		M/S Azam Trading Services	074601 Applied 09-11-2017	JS Cath IV Catheter, Dukwoo Medical Co Ltd Korea	Selected
112	3 way stop cock (without extension)	M/S Three13 Pharma	MDIR-0000229 2021	KAWA	Selected
		M/S Sanwa Trading Corporation	Not Given	China	Rejected Poor quality
		M/S Meher Traders	Applied 07-06-2021	Wellmed, Suzhou Health Plastic Products China	Selected
		M/S Farooq Agencies	Applied 13-09-2019	Safety	Selected
		M/S Usman Co International	MDIE-0000164	Three Way Stop cock Plain, Bicakcilar Dis Ticaret A.S. Turkey	Selected
		M/S Imran Traders	MDMR-00079 15-09-2020	Safeone, Silver Surgical Complex Pakistan	Selected
		M/S Azam Trading Services	MDIR-0000876 10-10-2019	Discifix 3 Way stopper without extension, B Braun Pakistan Pvt Ltd	Selected

113	3 way stop cock (with extension)	M/S Three13 Pharma	MDIR-0000229 2021	KAWA	Selected
		M/S Sanwa Trading Corporation	Not Given	China	Rejected Poor quality
		M/S Meher Traders	Applied 07-06-2021	Wellmed, Suzhou Health Plastic Products China	Selected
		M/S Farooq Agencies	Applied 13-09-2019	Safety	Selected
		M/S Usman Co International	MDIE-0000164	Three Way Stop cock with 10cm Tubing, Bicakilar Dis Ticaret A S, Turkey	Selected
		M/S Azam Trading Services	MDIR-0000876 10-10-2019	Discofix J Way stopper with extension, B Braun Pakistan Pvt Ltd	Selected
114	CVP monitoring line	M/S Safeway Systems Pakistan	MDIR-0002845 10-09-2021	Welford Medical M Sdn Bhd Malaysia	Selected
		M/S Usman Co International	MDIE-0000161	Extension Line-250 PSI, Bicakilar Dis Ticaret A S, Turkey	Rejected Wide bore
		M/S Cardiac Care	0001650 12-085-2020	Vygon, Europe	Selected
115	Oropharyngeal airway (Adult)	M/S Sadqain Healthcare (Pvt) Ltd	MDIE-0000060 2020	One-piece Guedel airway, Intersurgical	Selected
		M/S Cardiac Care	Not Registered	Pahsco, Taiwan	Selected
		M/S Hakimsons (Pvt) Ltd	Applied 03-07-2020	Foyomed, Ningbo MFLAB Medical/ China	Selected
116	Oropharyngeal airway (Peads)	M/S Sadqain Healthcare (Pvt) Ltd	MDIE-0000060 2020	One-piece Guedel airway, Intersurgical	Selected
		M/S Cardiac Care	Not Registered	Pahsco, Taiwan	Selected
		M/S Hakimsons (Pvt) Ltd	Applied 03-07-2020	Foyomed, Ningbo MFLAB Medical/ China	Selected
117	Urine bag (Adult)	M/S Sanwa Trading Corporation	Not Given	China	Selected
		M/S Meher Traders	Applied 22-11-2019	Wellmed, Changzhou Yaping Medical China	Selected
		M/S Farooq Agencies	MDIE-0000151 28-10-2020	Star	Selected
118	Urine bag (Peads)				
119	Urometer	M/S Surgimedika Pharma	Not Applied	Great Uro 2600ml Drain bag with 500ML meter, Ningbo Greentcare China	Selected
		M/S F.S Pharma	Applied	Hitecare China	Rejected Short length
		M/S Flowtronix Systems	Not Applied	China	Rejected
		M/S Meher Traders	Applied 22-11-2019	Wellmed, Changzhou Yaping Medical China	Rejected Leakage
		M/S Cardiac Care	Not Applied	Pahsco, Taiwan	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected Short length

120	NG tube all sizes (Adult)	M/S Samwa Trading Corporation	Not Given	China	Rejected Too soft
		M/S Meher Traders	Applied 22-11-2019	Wellmed, Changzhou Yaping Medical China	Rejected SNP
		M/S Farooq Agencies	Applied 07-09-2020	PIGEON	Rejected Tip connector too soft
		M/S Usman Co International	MDIR-0001765	Gastrogastric Catheter-Lavin (Radiopaque Line & DEHP Free), Bicakclar Dis Ticaret A.S Turkey	Selected
		M/S Hakimsons (Pvt) Ltd	MDIR-0003330 15-04-2022	Foyomed, Ningbo MFLAB Medical/ China	Selected
121	NG tube all sizes (Peads)	M/S Samwa Trading Corporation	Not Given	China	Rejected SNP
		M/S Meher Traders	Applied 22-11-2019	Wellmed, Changzhou Yaping Medical China	Rejected Poor quality
		M/S Farooq Agencies	Applied 07-09-2020	PIGEON	Selected
		M/S Usman Co International	MDIR-0003046	Feeding Tube (Radiopaque Line & DEHP Free), Bicakclar Dis Ticaret A.S, Turkey	Selected
		M/S Hakimsons (Pvt) Ltd	MDIR-0003330 15-04-2022	Foyomed, Ningbo MFLAB Medical/ China	Selected
122	ECG electrodes (Adult)	M/S FW Distributor	Applied 05-04-2021	TOPTRACE- Ceracarta, Italy	Selected
		M/S Meher Traders	Applied 06-05-2021	Wellmed, China Qingdao Bright Medical Manufacturing	Rejected Poor quality adhesive
		M/S Farooq Agencies	MDIE-0000372 22-02-2022	Safety	Selected
		M/S Cardiac Care	MDIE-0000292 15-07-2021	Fiab, Italy	Rejected Poor quality
		M/S Hakimsons (Pvt) Ltd		My Heart, Qingdao Bright Medical/ China	Selected
		M/S Magma	Applied	ECG Electrode Red Dot 2228, 3M, USA	Selected
123	ECG electrodes (Peads)	M/S FW Distributor	Applied 05-04-2021	TOPTRACE- Ceracarta, Italy	Selected
		M/S Meher Traders	Applied 06-05-2021	Wellmed, China Qingdao Bright Medical Manufacturing	Rejected Poor quality adhesive
		M/S Cardiac Care	MDIE-0000292 15-07-2021	Fiab, Italy	Rejected SNP
		M/S Hakimsons (Pvt) Ltd		Swaromed, Nissha Medical Technologies Ltd, UK	Selected
		M/S Magma	Applied	ECG Electrode Red Dot 2228, 3M, USA	Selected
124	Suction catheter All Sizes (Adult)	M/S Meher Traders	Applied 22-11-2019	Wellmed, Changzhou Yaping Medical China	Selected
		M/S Farooq Agencies	Applied 07-09-2020	PIGEON	Selected
		M/S Hakimsons (Pvt) Ltd	MDIR-0003326 15-04-2022	Foyomed, Ningbo MFLAB Medical/ China	Selected

125	Suction catheter All Sizes (Pends)	M/S Meher Traders	Applied 22-11-2019	Wellmed, Chonghua Yaping Medical China	Selected
		M/S Farooq Agencies	Applied 07-09-2020	PIGEON	Selected
		M/S Hakimsons (Pvt) Ltd	MDIR-0003326 15-04-2022	Foyumed, Ningbo MFLAB Medical/ China	Selected
126	IV Burette	M/S Three13 Pharma	MDIR-0002285 2015	UNISON	Selected
		M/S Meher Traders	MDIR-0002625	Wellmed, Hunan Luzhou China	Rejected can't be refilled once disc is closed too soft
		M/S Farooq Agencies	063914 09-07-2010	Golden	Selected
		M/S Safeway Systems Pakistan	MDIR-0002843 10-09-2021	Welford Medical M Sdn Bhd Malaysia	Rejected Dose closes refilling not possible
		M/S Hakimsons (Pvt) Ltd	MDIR-0003328 15-04-2022	MedSuyun, Jiangsu Suyun Medical, China	Rejected Disc lock can't be refilled
127	IV set	M/S Three13 Pharma	MDIR-059236 2019	Shifa	Selected
		M/S Nisa SF (Pvt) Ltd	MDMR-000011	BM Disposable IV Set Blister pack, Nisa SF	Rejected Obviously poor manufacturing quality
		M/S Meher Traders	MDIR-0002368 29-03-2021	Wellmed, Hunan Luzhou China	Rejected can't be refilled once disc is closed too soft
		M/S Farooq Agencies	071664	Trust (Blister)	Rejected Poor quality manufacturing
		M/S Azam Trading Services	MDIR-0001368 26-02-2020	Starle Flo IV Set Basic, Jiangsu Shenli Medical Production Co. Ltd, China	Rejected SNP
128	Disposable syringes 1ml	M/S Three13 Pharma	059233 2019	Shifa	Selected
		M/S Nisa Impex (Pvt) Ltd	MDIR-0002878	Nisa Impex	Selected
		M/S Amson Vaccines & Pharma	MDIR-000114 10-03-2021	Apple K1, Amson Vaccines & Pharma	Rejected 3ml size is labeled as 1ml
		M/S Global Health Care	MDIR-0002319 2021	TroJecor U-100, Troge Medical Germany	Selected
		M/S Farooq Agencies	059207 17-11-2009	Star	Rejected Needle attached in built needed
129	Disposable syringes 3ml	M/S Three13 Pharma	059295 2019	Shifa	Selected
		M/S Nisa Impex (Pvt) Ltd	MDIR-0001282	Nisa Impex	Rejected Plunger gives way
		M/S Nisa SF (Pvt) Ltd	MDMR-000008	BM Disposable Auto Disable Syringe 3ml Blister pack, Nisa SF	Rejected Poor quality
		M/S Amson Vaccines & Pharma	MDIR-00031 07-12-2018	Apple K1, Amson Vaccines & Pharma	Selected
		M/S Meher Traders	MDIR-0002550 26-05-2021	Wellmed, Hunan Luzhou China	Selected
		M/S Farooq Agencies	MDMR-000150 16-09-2021	Trust	Rejected Plunger gives way
		M/S Usman Co International	MDIR-0001761	Yixin Auto Disable, Jiangsu Sanxin Medtec Co. Ltd, China	Rejected Luer lock required

130	Disposable syringes 5ml	M/S Three13 Pharma	059295 2019	Shifa	Selected
		M/S Nisa Impex (Pvt) Ltd	MDIR-0001282	Nisa Impex	Rejected Plunger gives way
		M/S Nisa SF (Pvt) Ltd	MDMR-000008	BM Disposable Auto Enable Syringe 5cc with needle blister pack, Nisa SF	Rejected Very good quality, test - does not fit with plunger Plunger leaves
		M/S Amson Vaccines & Pharma	MDIR-000132 07-12-2018	Apple K1, Amson Vaccines & Pharma	Selected
		M/S Meher Traders	MDIR-0002550 26-05-2021	Wellmed, Hunan Luzhou China	Selected
		M/S Farooq Agencies	MDMR-000150 16-09-2021	Trust	Rejected Plunger gives way
		M/S Usman Co International	MDIR-0001761	Yixin Auto Enable, Jiangxi Sanxin Medtec Co Ltd, China	Rejected Lear lock required
		M/S Imran Traders	MDMR-00078 12-08-2020	Orange, Silver Surgical complex Pakistan	Selected
131	Disposable syringes 10ml	M/S Three13 Pharma	059234 2019	Shifa	Selected
		M/S Nisa SF (Pvt) Ltd	MDMR-000009	BM Disposable Syringe 10cc with needle blister pack, Nisa SF	Rejected Plunger gives way to blood
		M/S Amson Vaccines & Pharma	MDIR-000149 07-12-2018	Apple, Amson Vaccines & Pharma	Selected
		M/S Farooq Agencies	071663	Trust	Rejected Poor quality manufacturing
		M/S Usman Co International	MDIR-0002761	UCI Disposable Syringe, Jiangxi Sanxin Medtec Co Ltd, China	Selected
132	Disposable syringes 20ml	M/S Three13 Pharma	059234 2019	Shifa	Rejected Lear lock needed
		M/S Nisa SF (Pvt) Ltd	MDMR-000009	BM Disposable Syringe 20ml blister pack, Nisa SF	Rejected Plunger gives way to blood
		M/S Farooq Agencies	059207 17-11-2009	Star	Rejected Lear lock required
		M/S Usman Co International	MDIR-0002761	UCI Disposable Syringe, Jiangxi Sanxin Medtec Co Ltd, China	Rejected Lear lock required
133	Disposable syringes 50ml	M/S Three13 Pharma	059235 2019	Shifa	Selected
		M/S Nisa Impex (Pvt) Ltd	MDIR-0002879	Nisa Impex	Selected
		M/S Farooq Agencies	059246 29-12-2009	Star	Rejected Lear lock required
		M/S Usman Co International	MDIR-0002761	UCI Disposable Syringe, Jiangxi Sanxin Medtec Co Ltd, China	Selected
134	Disposable syringes 60ml	M/S Three13 Pharma	059235 2019	Shifa	Selected
		M/S Nisa Impex (Pvt) Ltd	MDIR-0002879	Nisa Impex	Selected
		M/S Farooq Agencies	059246 29-12-2009	Star	Selected

135	Tegaderm dressing large Or equivalent	M/S Threel3 Pharma	MDIE-0000176 2021	OXYMAX	Rejected Poor adhesive
		M/S Cardiac Care	Not Applied	TPRI, China	Selected
		M/S Ali & Sons Enterprises	Applied April 2022	Alban, Ningbo Chinmed Technology Co Ltd	Selected
		M/S Magma	Applied	Tegaderm 1626, 3M, USA	Selected
136	Transparent surgical tape transparent water resistant medical grade adhesive latex free and hypoallergic	M/S Threel3 Pharma	MDIE-0000179 2021	Surgitrans	Rejected Least adhesive does not cling to skin at all
		M/S Farooq Agencies	Applied 02-08-2019	RECO 3x10	Rejected Poor adhesive
		M/S Ali & Sons Enterprises	Applied April 2022	Alban, Ningbo Chinmed Technology Co Ltd	Rejected SNP
		M/S Magma	Applied	Transpore Tape 1527-1, 1527-2, 1527-3, 3M, USA	Selected
137	Corafix tape Or equivalent	M/S Imran Traders	Applied	Protect Fix, Zhejiang Hongyu China	Selected
138	External defib pads (Adult)	M/S Cardiac Care	Applied Feb-2021	Fiab, Italy	Selected
		M/S A & E Medical	Not Applied	Heart Sync CE Mark, NISSHA Medical Technologies, UK/USA	Selected
139	External defib pads (Peads)	M/S Cardiac Care	Applied Feb-2021	Fiab, Italy	Rejected SNP
140	CVP line	M/S UDL Distribution (Pvt) Ltd	MDIR-0000437 19-06-2019	Arrow, USA (Teleflex)	Selected
		M/S Flowtronix Systems	MDIR-0001067 13-12-2019	ABLE, Guangdong Baihe Medical, China	Rejected
141	Breathing circuit with bag (Adult)	M/S 4A International	MDIR-0002307 2021	EM05-003C-1, Excellentcare Medical, China	Selected
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001313 2020	22mm Compact, extendable breathing system with 2L bag and 1.5m limb, Intersurgical	Selected
		M/S Meher Traders	MDIR-0002375 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Rejected Needs light TV as not collapsible
		M/S Cardiac Care	Applied 02-03-2020	Hsiner, Taiwan	Rejected Poor quality
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected Collapsible circuit required
142	Breathing circuit with bag (Peads)	M/S 4A International	MDIR-0002307 2021	EM05-004B, Excellentcare Medical, China	Selected
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001313 2020	15mm Compact, paediatric extendable breathing system with luer elbow, Intersurgical	Selected
		M/S Cardiac Care	Applied 02-03-2020	Hsiner, Taiwan	Rejected SNP
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected Collapsible circuit required

Handwritten signatures and initials are present at the bottom of the page, including a circled number '59'.

143	Cotton bandage	M/S Cotton Craft (Pvt) Ltd	MDMR-006272	Local Manufacturer	Selected
		M/S Karim Industries	MDME-0000030 25-02-2021	Medi Band, Karim Industries	Selected
144	Yankuer suction with Handel	M/S Freesia Enterprises	Applied 2019	Canack	Selected
		M/S Meher Traders	Applied 02-12-2020	Wellmed, Ningbo Greatmountain Medical China	Rejected Hole not required on top
		M/S Usman Co International	MDIR-0003047	Yankuer Suction Connecting Set + Handle Bulb Tip Cut to Fit a 300cm, Bicakcilar Dis Ticaret A S Turkey	Rejected Narrow diameter
		M/S Cardiac Care	Applied June-2021	Greenmed, China	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected SNP
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Cardiac Surgery Disposable Cannula, Bismil International Pakistan	Rejected Short length
145	IV injector / pressure line 200ml <i>Coil Lah Islam</i>	M/S Safeway Systems Pakistan	MDIR-0002845 10-09-2021	Welford Medical M Sdn Bhd Malaysia	
		M/S Usman Co International	MDIE-0000165	Pressure Line 800 PSI M/F/Length 200cm, Bicakcilar Dis Ticaret A S, Turkey	Selected
		M/S Cardiac Care	MDIR-0001646 12-08-2020	Pro-Active, Italy	Selected
146	Cath mount	M/S 4A International	MDIR-0002303 2021	EM07-004A, Excellentcare Medical, China	Selected
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001157 2020	Flexible fixed elbow catheter mount 22F-flip top cap with 7 mm port-2, Intersurgical	Selected
		M/S Meher Traders	Not Applied	Wellmed, Ningbo Greatmountain Medical China	Rejected Side port difficult to place back
		M/S Cardiac Care	Applied 06-03-2020	Hsiner, Taiwan	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Rejected Dead space too large
147	Folleys silicon catheters	M/S FW Distributor	Applied 11-09-2021	Folysil-Coloplast, Denmark	Selected
		M/S Meher Traders	MDIR-0002860 03-12-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Cardiac Care	Not Registered	Puhco, Taiwan	Selected



148	Nebs mask (Adult)	M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001317 2020	Adult, aerosol mask with nose clip, Intersurgical	Selected
		M/S Sanwa Trading Corporation	Not Given	China	Selected
		M/S Meher Traders	MDIR-0002372 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	MDIR-0001789 16-09-2020	Safety	Selected
		M/S Cardiac Care	Applied 02-01-2020	Hsiner, Taiwan	Selected
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Selected
149	Nebs mask (Peads)	M/S Meher Traders	MDIR-0002372 29-03-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	MDIR-0001789 16-09-2020	Safety	Selected
		M/S Cardiac Care	Applied 02-01-2020	Hsiner, Taiwan	Rejected SNP
		M/S Hakimsons (Pvt) Ltd		Foyomed, Ningbo MFLAB Medical/ China	Selected
150	Nebs mask (Neonate)	M/S Farooq Agencies	MDIR-0001789 16-09-2020	Safety	Selected
151	T piece nebs	M/S 4A International	MDIR-0002300 2021	EM06-003T, Excellentcare Medical, China	Rejected Nozzle too hard to attached with flowmeter breaks
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001756 2020	Cirrus nebuliser breathing system T-Kit, 22mm, Intersurgical	Selected
152	Dialysis catheter	M/S UDL Distribution (Pvt) Ltd	MDIR-0000428 31-05-2019	Arrow, USA (Teleflex)	Selected
		M/S Economic Health Enterprises	Not Applied	Smith's Medical International	Rejected Not registered
		M/S Iqbal & Company	MDIR-0000611 2019	Double Lumen Dialysis Catheter, Medcomp, USA	Selected
		M/S Cardiac Care	Applied Nov-2021	Kinval, UK	Rejected Not registered
153	Peritoneal Dialysis Catheter				

154	Hemodialysis Complete kit	M/S Eastern Medical Care (Pvt) Ltd	MDIR-0001597 (2020) MDIR-0002346 (2021) MDIR-0001595 (2020) MDIR-0002366 (2021) MDIR-0001596 (2020) MDIR-0002647 (2021) MDIR-0002649 (2021)	Dialyzer (Nipro Corporation Japan/China) Blood Tubing Lines (Nipro Corporation Indonesia) Fistula Needles (Nipro Corporation Indonesia/Thailand) Farmacart B-bag (Farmasol Tibbi, Turkey)	Selected
		M/S Iqbal & Company	Polyflux Dialyzer 14L/17L, 1.4 m <sup>2</sup> /1.7m <sup>2</sup> , Steam Sterilized, 50 micron, Baxter, Gambro Germany (Applied Dec 2021) Bicarb Solution with Powder (A+B) 4 Ltr can Jassh Pharma Lahore (Applied Dec-2021)	Blood Line with Fluid Barrier, Taiwan/ china (Not Applied) A.V Fistula Needle 16G/17G Taiwan/China (Not Applied)	Selected
155	Cotton roll	M/S Cotton Craft (Pvt) Ltd	MDMR-006271	Local Manufacturer	Selected
		M/S Kohinoor Industries			Selected
		M/S Medicare Enterprises	052655 2008	Local Manufacturer, Surgee Wool, Surgical Fibre	Selected
		M/S Karim Industries	MDME-0000037 2021	Medi Cost, Karim Industries	Selected
156	Sterilized Surgical gauze	M/S Threel3 Pharma	MDIR-030772	Surgigauz	Selected
		M/S Cotton Craft (Pvt) Ltd	MDMR-000123	Local Manufacturer	Selected
		M/S Kohinoor Industries			Selected
		M/S Karim Industries	050293 28-06-2008	Soft Gauze, Karim Industries	Rejected Poor quality
157	Sterilized Gauze Roll	M/S Cotton Craft (Pvt) Ltd	MDMR-006269	Local Manufacturer	Selected
		M/S Kohinoor Industries			Selected
		M/S Medicare Enterprises	066204 2010	Local Manufacturer, Hospitex Gauze, General Pharma	Rejected Poor quality
		M/S Karim Industries	008396 15-05-1985	Absorbent Gauze, Karim Industries	Selected
158	Dignity sheets	M/S Sanwa Trading Corporation	Not Given	China	Rejected Poor quality Raja bazar brand
		M/S Farooq Agencies	Applied 13-01-2021	Star	Selected
		M/S Imran Traders	Applied	National, China	Selected

*[Handwritten signature]*

*[Handwritten signature]*


*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

(3)

159	Crape bandage	M/S Cotton Craft (Pvt) Ltd	MIDMR-008373	Local Manufacturer	Selected
		M/S Kohinoor Industries			Selected
		M/S Medicare Enterprises	066205 2010	Local Manufacturer, Hospital Crep, General Pharma	Selected
		M/S Karim Industries	052592 26-09-2008	Medi Crep Bandage, Karim Industries	Selected
160	ACT pipettes	M/S Iqbal & Company	MDIR-0002261 2021	ACT Cartridge, Accriva/Werfen II., ITC, USA	Selected
		M/S Safeway Systems Pakistan	MDIR-0002922 03-11-2021	Sienco Inc USA	Rejected Handling too difficult
		M/S Medtronic Pakistan	MDIR-0000157 2019	IIR-ACT, Medtronic Minneapolis, MN USA	Rejected Too difficult
161	Oxygen Therapy mask with integrated Filter & Tub (Filt Mask or Equivalent)	M/S Sadqain Healthcare (Pvt) Ltd		FitaMask, Eco, adult, medium concentration mask with tube, 2.1m, Intersurgical	Rejected
		M/S Meher Traders	Not Applied	Inter Surgical, UK	Rejected SNP
162	Examination gloves	M/S Moon Enterprises	attached 2022	AXG Industries SDN BHD, Malaysia	Selected
		M/S Economic Health Enterprises	Not Applied	Top Gloves, Malaysia	Selected
		M/S Eastern Medical Care (Pvt) Ltd	MDIE-000251,252,253 2021	ASAP International Sdm. Bhd, Malaysia	Rejected Poor Quality
		M/S Freesia Enterprises	Applied 2021	Safchand	Selected
		M/S Sanwa Trading Corporation	Not Given	Qualatex, Malaysia	Rejected Poor quality
		M/S Farooq Agencies	Applied 13-02-2022	QC Premium	Selected
		M/S Hope Pharma	Applied 27-07-2020	Sempermed, Semperit Asis Pte Ltd, Austria/ Malaysia	Selected
		M/S AL-Hamd Enterprises	MDIE-0000228	Optishield, Tan Sin Lian Industries Sdn.Bhd Malaysia	Selected
		M/S Safeway Systems Pakistan	Not Applied	Vogt Medical GmbH Germany	Rejected
		M/S Endonald Biomedica	Not Applied	Rayzen, Rayzen Tainjin Health Care, China	Selected



163	Surgical gloves	M/S Three13 Pharma	MDIR-0001623 2020	Maxitex	Selected
		M/S Freesia Enterprises	Applied 2019	Comfit Premium	Selected
		M/S Sanwa Trading Corporation	Not Given	Ying Med, China	Rejected Rough size bigger
		M/S Farooq Agencies	MDIR-0001307 18-02-2022	Medispo	Rejected loose/ poor quality
		M/S AL-Hamd Enterprises	MDIR-0001236	Surgitex, Suzhou Colour Way New Material Co. Ltd China	Selected
		M/S Safeway Systems Pakistan	Not Applied	Vogt Medical GmbH Germany	Rejected loose size poor quality
		M/S Endoaid Biomedica	Not Applied	Raysen, Raysen Tainjin Health Care, China	Rejected Rough finishing
		M/S Kaumedex	MDIR-0001235 2020	HighMax, Supermax Gloves Manufacturing Sdn. Bhd. Malaysia	Selected
164	Disposable Gown reinforced	M/S Flowtronix Systems	Not Applied	EN, Elate C.C (Pvt) Ltd, Pakistan	Selected
		M/S Farooq Agencies	MDIE-0000102 20-08-2020	National	Rejected Poor quality
		M/S Safeway Systems Pakistan	Not Applied	Starset Startip Turkey	Rejected SNP
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Disposable Surgical Reinforce Gown Large, Skybro Heal Care Pakistan	Rejected
		M/S Magma	Applied	Disposable Reinforced Gown, Medpro, Pakistan	Selected
165	Water Resistant Gown (60) gsm	M/S Flowtronix Systems	Not Applied	EN, Elate C.C (Pvt) Ltd, Pakistan	Selected
		M/S Endoaid Biomedica	Not Applied	Disposable Gowns, Raysen Tainjin Health Care, China	Rejected
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Full Reinforced Disposable Surgical Gown, Skybro Heal Care Pakistan	Selected
		M/S Magma	Applied	Standard Surgical Gown Unsterilized, Medpro, Pakistan	Selected
166	Disposable Cap (Male)	M/S Three13 Pharma	Not Applied	China/ LOCL	Rejected SNP
		M/S Meher Traders	MDIE-000267 26-05-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	MDIE-0000098 20-08-2020	National	Rejected Poor quality
		M/S Endoaid Biomedica	Not Applied	Disposable Gowns, Raysen Tainjin Health Care, China	Rejected SNP
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Disposable Cap, Skybro Heal Care Pakistan	Selected

*Handwritten signature*

*Handwritten signatures and initials*

34

167	Disposable Cap (Female)	M/S Three13 Pharma	Not Applied	China/ LOCL	Selected
		M/S Sanwa Trading Corporation	Not Given	China	Selected
		M/S Meher Traders	MDIE-000267 26-05-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Disposable Cap, Skybro Heal Care Pakistan	Rejected Small
168	Disposable Shoe Cover	M/S Three13 Pharma	Not Applied	China/ LOCL	Rejected Poor quality
		M/S Sanwa Trading Corporation	Not Given	China	Selected
		M/S Meher Traders	MDIE-000268 26-05-2021	Wellmed, Ningbo Greatmountain Medical China	Selected
		M/S Farooq Agencies	MDIE-0000100 20-08-2020	National	Selected
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Disposable Shoe Cover, Skybro Heal Care Pakistan	Rejected Poor quality
169	Surgical Face Mask	M/S Three13 Pharma	Not Applied	China/ LOCL	Selected
		M/S Taha Enterprises	Not Applied	Executive Brand by ST Engineering Services	Rejected Poor quality
		M/S Meher Traders	MDIE-000269 26-05-2021	Wellmed, Ningbo Greatmountain Medical China	Rejected Quality issue
		M/S Farooq Agencies	MDIE-0000104 20-08-2020	National	Selected
		M/S Cardiac Care	Not Applied	Greetmed, China	Selected
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Disposable Face Mask, Skybro Heal Care Pakistan	Selected
		M/S Magma	Applied	Disposable Face Mask, Medpro, Pakistan	Selected
170	Surgical Face Mask Tie on	M/S Meher Traders	MDIE-000269 26-05-2021	Wellmed, Ningbo Greatmountain Medical China	Rejected Quality issue
		M/S Cardiac Care	Not Applied	Greetmed, China	Selected
		M/S M. Yousaf & Co	Applied 30-04-2021	Skymax Surgical Face Mask, Skybro Heal Care Pakistan	Selected
		M/S Magma	Applied	Disposable Face Mask Tie on, Medpro, Pakistan	Selected

Handwritten signatures and initials are present below the table. A circled number '33' is visible in the bottom right corner.


**DISINFECTANT FOR HOSPITAL**


171	High level disinfectant for OT Instruments (Ortho-phthalaldehyde (OPA) 0.55%)	M/S Economic Health Enterprises	Not Applied	OneSept PA, Turkey	Selected
		M/S Three13 Pharma	Not Applied	TRIACID N	Rejected Formula different ingredients are in low level disinfectant
		M/S 4A International	MDIR-0002559 2021	Detro OPA 5L., Detrol Healthcare, Turkey	Selected
		M/S Muller & Phipps (Pvt) Ltd	Applied	Johnson & Johnson, ASP, Switzerland	Selected
		M/S Global Health Care	Not Applied	I&D Hydroxy PPA (Surgical Instrument & Endoscope), DTI Turkey	Selected
		M/S Ali & Sons Enterprises	Applied Dec-2020	Persafe, Lanxess/Antech Corp UK	Rejected Formula different ingredients are in low level disinfectant
172	Hand Disinfectant/Sanitizer (60-90% Alcohol (Ethyl or Isopropanol) with moisturizer)	M/S Economic Health Enterprises	PSQA CM/L-4347-2020	Surgillum, Nigehban/ Promark Local	Selected (Subject to provision of AFIP report which contain: 1. Time to kill. 2. Efficacy against MRSA, VRE & ESKAPE. 3. MIC of disinfectant against MRSA, VRE & Acinetobacter
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001448 2020		
		M/S WAH Nobel Acetates Ltd	CML/PZ 296/2021	Local Manufacturer (Wah Nobel Acetates Ltd)	
		M/S Next Pharmaceuticals Product	2997720 2020	Nextol Hand Disinfectant/Sanitizer Ethyl Alcohol Liquid/Gel 500ml with pump	
		M/S Global Health Care	MDMR-000058 2020	Hiclean Alcoholic, Medinostic	
173	Skin Disinfectant (Methylated Spirit)	M/S Three13 Pharma	Not Applied	MAYO/3STR	Selected
174	Soda Lime (Pack of 5 liters)	M/S 4A International	MDIR-0002306 2021	EM09-001, Excellence Medical, China	Selected
		M/S Sadqain Healthcare (Pvt) Ltd	MDIR-0001156 2020	Intersorb Plus jerican, pink to white color change, 5L, Intersurgical	Rejected SNP
175	Topical Antiseptic	M/S Brooks Pharma (Pvt) Ltd	009528 July-2018	Povidone Iodine 10% Solution, Pyodine Solution 450ml Brooks Pharma Pakistan	Selected
176	Topical Antiseptic Scrub (Povidone-iodine Topical Solution (10% W/V))	M/S Global Health Care		Hiclean CHX, Medinostic	Rejected Formula different
		M/S Brooks Pharma (Pvt) Ltd	010403 Aug-2018	Pyodine Surgical Scrub 450ml, Povidone Iodine 7.5% Scrub, Brooks Pharma Pakistan	Selected
177	Topical/ Skin Scrub (Alcohol 60-90% (Ethyl or Isopropanol) + Chlorhexidine gluconate 1% with moisturizer )	M/S Global Health Care	MDMR-000050 2020	Hiclean CHX, Medinostic	Rejected Formula different
		M/S Magma	MDMR-00436.001 2017	Avagard 9200 500ml, 3M, USA	Rejected Formula different


178	Surface Disinfectant (Pentapotassium/ Benzenesulfonic Acid/ Formic Acid / Malic Acid/Sulphamic Acid/ Chlorine Based)	M/S Economic Health Enterprises	Not Applied	Virusolve+, Amity International UK	Rejected Formula different
		M/S Three13 Pharma	Not Applied	DESCOCID N	Rejected Formula different
		M/S 4A International	MDIR-0002557 2021	Deiro San Floor II, Deiro Healthcare, Turkey	Rejected Formula different
		M/S Global Health Care	MDIR-0000088 2019	Chemegone, Melmark UK	Rejected Formula different
		M/S Ali & Sons Enterprises	Not Applied	Virkon, Lantess/Anstoch Corp UK	Selected
<b>CENTRAL PURCHASE (Quote these items in separate envelope)</b>					
179	Oxygenator Adult coated with custom pack and arterial filter with stand (FDA Approved)	M/S Cardiac Care	MDIR-0001342 18-02-2020	Euroset, Italy	Selected
		M/S Medtronic Pakistan	MDIR-0000176 2019	Affinity Fusion, Medtronic Minneapolis, MN USA	Selected
		M/S Azam Trading Services	MDIR-0002552 26-05-2021	Capiox Oxygenator, Terumo Cardiovascular System Corporation	Selected
180	Oxygenator small Adult coated with custom pack and arterial filter with stand (FDA Approved)	M/S Medtronic Pakistan	MDIR-0000176 2019	Affinity Fusion, Medtronic Minneapolis, MN USA	Selected
		M/S Azam Trading Services	MDIR-0002552 26-05-2021	Capiox Oxygenator, Terumo Cardiovascular System Corporation	Selected
181	Oxygenator Peads coated with custom pack and arterial filter with stand (FDA Approved)	M/S Cardiac Care	MDIR-0001343 18-02-2020	Euroset, Italy	Selected
		M/S Medtronic Pakistan	MDIR-0002032 2020	Affinity Pixie, Medtronic Minneapolis, MN USA	Selected
		M/S Azam Trading Services	MDIR-0002552 26-05-2021	Capiox Oxygenator, Terumo Cardiovascular System Corporation	Selected
182	Oxygenator Neonates coated with custom pack and arterial filter with stand (FDA Approved)	M/S Medtronic Pakistan	MDIR-0002032 2020	Affinity Pixie, Medtronic Minneapolis, MN USA	Selected
		M/S Azam Trading Services	MDIR-0002552 26-05-2021	Capiox Oxygenator, Terumo Cardiovascular System Corporation	Selected
183	Mitral Valve Mechanical All sizes Bi-Leaflet, made of pyrolite (FDA Approved)	M/S Verizon	MDIR-0000238 2019	Mitral Valve Mechanical Bi- leaflet, St. Jude Medical, USA	Selected
		M/S Cardiac Care	MDIR-0000734 27-09-2019	On-X Cryolife, USA	Selected
		M/S Medtronic Pakistan	MDIR-0001016 2019	Medtronic Open Pivot (Standard) Minneapolis, MN USA	Selected
184	Mitral Valve Biological Porcine All sizes (FDA Approved)	M/S Cardiac Care	MDIR-0000458 24-06-2019	On-X Cryolife, USA	Selected
		M/S Medtronic Pakistan	MDIR-0000159 2019	Hancock II Ultra Bioprostheses, Minneapolis, MN USA	Selected


37


185	Aortic Valve (Mechanical) Size 21 and above Bi-Leaflet, made of pyrolite Carbon (FDA Approved)	M/S Verizon	MDIR-0000236 2019	Aortic Valve Mechanical Pericardial, St. Jude Medical USA	Selected
		M/S Cardiac Care	Applied Feb-2021	On-X Cryolife, USA	Selected
		M/S Medtronic Pakistan	MDIR-0001014 2019	Medtronic Aortic Valve (Standard), Minneapolis, MN USA	Selected
186	Aortic Valve (Mechanical) Size 19 and below Bi-Leaflet, made of pyrolite Carbon, small / reduced swing cuff with low gradient for small annulus (FDA Approved)	M/S Verizon	MDIR-0000237 2019	Aortic Valve Mechanical Pericardial, St. Jude Medical USA	Selected
		M/S Medtronic Pakistan	MDIR-0001014, 1017 2019	Medtronic Aortic Valve (Standard), Minneapolis, MN USA	Selected
187	Aortic Valve (Biological) Bovine Pericardial All sizes (FDA Approved)	M/S Verizon	Applied 2020	Aortic Valve Bovine Pericardial, St. Jude Medical USA	Selected
		M/S Medtronic Pakistan	Not Applied	Aortic Valve Bovine Pericardial, Minneapolis, MN USA	Selected
	Aortic Valve (Biological) Porcine All sizes (FDA Approved)	M/S Cardiac Care	Applied Feb-2021	On-X Cryolife, USA	Selected
		M/S Medtronic Pakistan	MDIR-0000160 2019	Porcine Aortic Valve Pericardial, Minneapolis, MN USA	Selected

  
**Dr. Tauqeer Akbar**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi

  
**Dr. Musfereh Siddiqueh**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi


  
**Dr. Qudsia Anjum Quershi**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi

  
**Dr. Zainab Farid**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi

  
**Lt Col (R) Dr. Arif Maqsood Ali**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi

  
**M Noor ul Amin**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi

  
**Dr. Humaira Kiyani**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi

  
**Dr. Gul e Raana Waseem**  
 Member/ Secretary  
 Technical Advisory Committee  
 Rawalpindi

  
**Dr. Javed Iqbal**  
 End User (Member)  
 Technical Advisory Committee  
 Rawalpindi

